

# Go Direct<sup>®</sup> Electrode Amplifier

(Order Code GDX-EA)



Use our Go Direct Electrode Amplifier to collect data with almost any electrode that has a BNC connector. This includes electrodes from Vernier, such as our pH electrodes, ion-selective electrodes (ISE), or oxidation-reduction potential (ORP) electrode. The Go Direct Electrode Amplifier can produce an output in mV, pH (when using a pH electrode).

**Note:** Vernier products are designed for educational use. Our products are not designed nor are they recommended for any industrial, medical, or commercial process such as life support, patient diagnosis, control of a manufacturing process, or industrial testing of any kind.

## What's Included

- Go Direct Electrode Amplifier
- Micro USB Cable

## Compatible Software

See [www.vernier.com/manuals/gdx-ea](http://www.vernier.com/manuals/gdx-ea) for a list of software compatible with Go Direct Electrode Amplifier.

## Getting Started

Please see the following link for platform-specific connection information:

[www.vernier.com/start/gdx-ea](http://www.vernier.com/start/gdx-ea)

### Bluetooth Connection

1. Connect a BNC electrode to the amplifier.
2. Install Vernier Graphical Analysis<sup>®</sup> on your computer, Chromebook<sup>™</sup>, or mobile device. See [www.vernier.com/ga](http://www.vernier.com/ga) for software availability.
3. Charge your sensor for at least 2 hours before first use.
4. Turn on your sensor by pressing the power button once. The Bluetooth<sup>®</sup> LED will blink red.
5. Launch Graphical Analysis.
6. Click or tap Sensor Data Collection.
7. Click or tap your Go Direct sensor from the list of Discovered

### USB Connection

1. Connect a BNC electrode to the amplifier.
2. If using a computer or Chromebook, install Graphical Analysis. If using LabQuest, make sure LabQuest App is up to date. See [www.vernier.com/ga](http://www.vernier.com/ga) for Graphical Analysis availability or [www.vernier.com/downloads](http://www.vernier.com/downloads) to update LabQuest App.
3. Connect the sensor to the USB port.
4. Launch Graphical Analysis or turn on LabQuest. You are now ready to collect data.
5. This is a multi-channel sensor. To change the default channel selections, see [www.vernier.com/start/gdx-ea](http://www.vernier.com/start/gdx-ea)

Wireless Devices. Your sensor's ID is located near the barcode on the sensor. The Bluetooth LED will blink green when it is successfully connected.


**Note:** This sensor does not work with the original LabQuest. It works with LabQuest 2 or LabQuest 3.

8. This is a multi-channel sensor. The active channel is listed in the Connected Devices Sensor Channels list. To change channels, select the check box next to the Sensor Channel(s) you would like to activate.
9. Click or tap Done to enter data-collection mode.

## Charging the Sensor

Connect the Go Direct Electrode Amplifier to the included USB micro cable and any USB device for two hours.

You can also charge up to eight Go Direct Electrode Amplifiers using our Go Direct Charge Station, sold separately (order code: GDX-CRG).



Charging	Blue LED is on steady while sensor is connected to the Charging Cable or Charge Station.
Fully charged	Blue LED is off when charging is complete.

## Powering the Sensor

Turning on the sensor	Press button once. Red LED indicator flashes when unit is on.
Putting the sensor in sleep mode	Press and hold button for more than three seconds to put into sleep mode. Red LED indicator stops flashing when sleeping.

## Connecting the Sensor

See the following link for up-to-date connection information:

[www.vernier.com/start/gdx-ea](http://www.vernier.com/start/gdx-ea)

### Connecting via Bluetooth

Ready to connect	Red LED flashes when sensor is awake and ready to connect.
Connected	Green LED flashes when sensor is connected via Bluetooth.

### Connecting via USB

Connected and charging	Blue and Green LEDs are solid when sensor is connected to Graphical Analysis via USB and unit is charging. (Green LED is obscured by the blue one.)
Connected, fully charged	Green LED is solid when sensor is connected to Graphical Analysis via USB and the unit is fully charged.
Charging via USB, connected via Bluetooth	Blue LED is solid and green LED is flashing, but the green flashing LED looks white because it is overwhelmed by the blue.

## Using the Product

Connect the sensor following the steps in the Getting Started section of this user manual.

### Channels

Go Direct Electrode Amplifier has two measurement channels:

- Potential
- pH

### Calibration

If a pH electrode is attached, a two-point calibration is recommended for the most accurate results. It is a simple process that takes only a few minutes. For additional calibration information, see [www.vernier.com/til/4011](http://www.vernier.com/til/4011)

In order to calibrate a pH electrode/electrode amplifier combination, or to confirm that a saved pH calibration is accurate, you should have a supply of pH buffer solutions that cover the range of the pH values you will be measuring. For more information about pH buffer solutions, including recipes for preparation, see [www.vernier.com/til/3625](http://www.vernier.com/til/3625)

Once you have calibrated a Go Direct sensor, the calibration is automatically stored to the sensor and will be used each time you connect to your device. You

can always choose to restore factory defaults if you feel the custom calibration is invalid.

If an ORP or ion-selective electrode is connected and you are measuring potential, calibration is not required. Calibration is disabled in this mode to ensure the best results.

## Specifications

Units	pH or mV
Input range (mV)	±1000 mV
Input range (pH)	0 to 14 pH
Accuracy with new electrode (mV)	±20 mV
Accuracy with new electrode (pH)	±0.2
Resolution	0.01 mV
Connections	
Wireless	Bluetooth
Wired	USB

## Care and Maintenance

### Battery Information

Go Direct Electrode Amplifier contains a small lithium-ion battery in the handle. The system is designed to consume very little power and not put heavy demands on the battery. Although the battery is warranted for one year, the expected battery life should be several years. Replacement batteries are available from Vernier (order code: GDX-BAT-300).

### Storage and Maintenance

To store Go Direct Electrode Amplifier for extended periods of time, put the device in sleep mode by holding the power button down for at least three seconds. The red LED will stop flashing to show that the unit is in sleep mode. Over several months, the battery will discharge but will not be damaged. After such storage, charge the device for a few hours, and the unit will be ready to go.

Exposing the battery to temperatures over 35°C (95°F) will reduce its lifespan. If possible, store the device in an area that is not exposed to temperature extremes.

### Water Resistance

Go Direct Electrode Amplifier is not water resistant and should never be immersed in water.

If water gets into the device, immediately power the unit down (press and hold the power button for more than three seconds). Disconnect the sensor and charging cable, and remove the battery. Allow the device to dry thoroughly before

attempting to use the device again. Do not attempt to dry using an external heat source.

## How the Sensor Works

Go Direct Electrode Amplifier is a circuit that allows the voltage output of the electrode to be monitored. The electrode is connected to the BNC connector on one end of the amplifier.

Go Direct Electrode Amplifier will default to report values in units of pH. If an ORP or ISE electrode is attached, however, you can choose to read in units of mV. For additional information, see [www.vernier.com/til/3984](http://www.vernier.com/til/3984)

## Troubleshooting

For troubleshooting and FAQs, see [www.vernier.com/til/3858](http://www.vernier.com/til/3858)

## Repair Information

If you have followed the troubleshooting steps and are still having trouble with your Go Direct Electrode Amplifier, contact Vernier Technical Support at [support@vernier.com](mailto:support@vernier.com) or call 888-837-6437. Support specialists will work with you to determine if the unit needs to be sent in for repair. At that time, a Return Merchandise Authorization (RMA) number will be issued and instructions will be communicated on how to return the unit for repair.

## Accessories/Replacements

### Item

Go Direct pH BNC Electrode

Go Direct ORP BNC Electrode

Go Direct Flat pH BNC Electrode

Go Direct Glass-Body pH BNC Electrode

Go Direct Potassium Ion-Selective Electrode BNC

Go Direct Nitrate Ion-Selective Electrode BNC

Go Direct Ammonium Ion-Selective Electrode BNC

Go Direct Chloride Ion-Selective Electrode BNC

Go Direct Calcium Ion-Selective Electrode BNC

Micro USB Cable

Go Direct 300 mAh Replacement Battery

USB-C to Micro USB Cable

### Order Code

GDX-PH-BNC

GDX-ORP-BNC

GDX-FPH-BNC

GDX-GPH-BNC

GDX-K-BNC

GDX-NO3-BNC

GDX-NH4-BNC

GDX-CL-BNC

GDX-CA-BNC

CB-USB-MICRO

GDX-BAT-300

CB-USB-C-MICRO

## Warranty

Warranty information for this product can be found on the Support tab at [www.vernier.com/gdx-ea](http://www.vernier.com/gdx-ea)

General warranty information can be found at [www.vernier.com/warranty](http://www.vernier.com/warranty)

## Disposal

When disposing of this electronic product, do not treat it as household waste. Its disposal is subject to regulations that vary by country and region. This item should

be given to an applicable collection point for the recycling of electrical and electronic equipment. By ensuring that this product is disposed of correctly, you help prevent potential negative consequences on human health or on the environment. The recycling of materials will help to conserve natural resources. For more detailed information about recycling this product, contact your local city office or your disposal service.

Battery recycling information is available at [www.call2recycle.org](http://www.call2recycle.org)

Do not puncture or expose the battery to excessive heat or flame.



The symbol, shown here, indicates that this product must not be disposed of in a standard waste container.

## Federal Communication Commission Interference Statement

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

### FCC Caution

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) this device may not cause harmful interference and
- (2) this device must accept any interference received, including interference that may cause undesired operation

### RF Exposure Warning

The equipment complies with RF exposure limits set forth for an uncontrolled environment. The antenna(s) used for this transmitter must not be co-located or operating in conjunction with any other antenna or transmitter. You are cautioned that changes or modifications not expressly approved by the party responsible for compliance could void your authority to operate the equipment.

## IC Statement

This device complies with Industry Canada license-exempt RSS standard(s). Operation is subject to the following two conditions:

- (1) this device may not cause interference, and
- (2) this device must accept any interference, including interference that may cause undesired operation of the device.

**Industry Canada - Class B** This digital apparatus does not exceed the Class B limits for radio noise emissions from digital apparatus as set out in the interference-causing equipment standard entitled "Digital Apparatus," ICES-003 of Industry Canada. Operation is subject to the following two conditions: (1) this device may not cause interference, and

- (2) this device must accept any interference, including interference that may cause undesired operation of the device.

To reduce potential radio interference to other users, the antenna type and its gain should be so chosen that the equivalent isotropically radiated power (e.i.r.p.) is not more than that permitted for successful communication.

**RF exposure warning:** The equipment complies with RF exposure limits set forth for an uncontrolled environment. The antenna(s) used for this transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.

Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes :

- (1) l'appareil ne doit pas produire de brouillage, et
- (2) l'appareil doit accepter toute interférence radioélectrique, même si cela résulte à un brouillage susceptible d'en compromettre le fonctionnement.

Cet appareil numérique respecte les limites de bruits radioélectriques applicables aux appareils numériques de Classe B prescrites dans la norme sur le matériel interférant-brouilleur: "Appareils Numériques," NMB-003 édictée par industrie Canada. L'utilisation est soumise aux deux conditions suivantes:

- (1) cet appareil ne peut causer d'interférences, et
- (2) cet appareil doit accepter toutes interférences, y comprises celles susceptibles de provoquer un dysfonctionnement du dispositif.

Afin de réduire les interférences radio potentielles pour les autres utilisateurs, le type d'antenne et son gain doivent être choisis de telle façon que l'équivalent de puissance isotrope émise (e.i.r.p) n'est pas plus grand que celui permis pour une communication établie.

**Avertissement d'exposition RF:** L'équipement est conforme aux limites d'exposition aux RF établies pour un environnement non supervisé. L'antenne (s) utilisée pour ce transmetteur ne doit pas être jumelée ou fonctionner en conjonction avec toute autre antenne ou transmetteur.

**Note:** This product is a sensitive measurement device. For best results, use the cables that were provided. Keep the device away from electromagnetic noise sources, such as microwaves, monitors, electric motors, and appliances.



Vernier Science Education  
13979 SW Millikan Way • Beaverton, OR 97005-2886  
Toll Free (888) 837-6437 • (503) 277-2299 • Fax (503) 277-2440  
info@vernier.com • www.vernier.com

Rev. 2/15/2024

Go Direct, Vernier Graphical Analysis, LabQuest, and other marks shown are our trademarks or registered trademarks in the United States. All other marks not owned by us that appear herein are the property of their respective owners, who may or may not be affiliated with, connected to, or sponsored by us.

The Bluetooth® word mark and logos are registered trademarks owned by the Bluetooth SIG, Inc. and any use of such marks by Vernier Science Education is under license. Other trademarks and trade names are those of their respective owners.

